



October 02, 2023

Client: Rogers Co. RWD #7

20352 South 4230 Road

Claremore, OK 74017

Requested By: -



National Environmental Laboratory Accreditation Program ODEQ TNI Certified

Sample Project Name: SDWIS Analysis Monthly

Date Samples Received: September 25, 2023 Time: 13:56 sample temp upon arrival at lab = 24.30°C

Matrix: Drinking Water

Lab Log Numbers: **FI25051-01** **FI25051-02** **FI25051-03** **FI25051-04**
 FI25051-05 **FI25051-06** **FI25051-07** **FI25051-08**
 FI25051-09 **FI25051-10**

Work Order: FI25051

Report # FI25051-1002231026

EPA Lab ID#'s: **Stillwater OK00092** **Tulsa OK00983** **OKC OK00129** **ICR OK 001**

Oklahoma Certification: Stillwater NELAP WasteWater, ODEQ 8316/ Drinking Water, DEQ D9602
 NELAP Tulsa WasteWater, ODEQ 9905 / Drinking Water, DEQ D9901
 Oklahoma City NELAP WasteWater ODEQ 7202 / Drinking Water, DEQ D9937

Kansas Certification: Stillwater NELAP CERT # E-10219

Method Reference: 40 CFR 136, 141, and 261 Methods for Chemical Analysis of Water and Wastes EPA-600/4-79-020, March 1983. Test Methods for Evaluating Solid Wastes, SW-846, Final Update VI. Standard Methods 2005 (21st Edition), Standard Methods 2011 (22nd Edition), Standard Methods 2017 (23rd Edition) for the Examination of Water and Wastewater.

Analysis Reference: If qualifiers present in "Prep Info" or "Analysis Info", then analysis performed as follows: @= Tulsa Lab and * = OKC Lab. If no qualifiers present, then analysis performed at Stillwater Lab.

Accurate Environmental Laboratories certify that the test results performed at the Stillwater lab meet all requirements of NELAP. Any exceptions to this can be found in the report notes, Quality Control section, or Method/Parameter section of the report.

- No cert. = Laboratory does not carry certification for this method/analysis.
- Non-TNI = Laboratory has state certification but method does not fall under TNI certification.

This report is to only be replicated in its entirety.

Revised or Amended reports supersede all previous reports.

Accurate Environmental sampling protocol was followed for any sampling performed by Accurate Field Services.

Field accreditation certification only applies to wastewater analysis. Field analysis is performed based on lab accreditation for appropriate drinking water methods until field certification program expands to include drinking water analysis.

Sample: LC030

Location Code: LC030

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 6:30

Lab Log# FI25051-01

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.0106 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:09 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:09 @PD

Sample: LC036

Location Code: LC036

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 7:00

Lab Log# FI25051-02

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.146 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:14 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:14 @PD

Sample: LC035

Location Code: LC035

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 6:30

Lab Log# FI25051-03

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.0450 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:18 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:18 @PD

Sample: LC047

Location Code: LC047

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 7:00

Lab Log# FI25051-04

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.0442 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:22 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:22 @PD

Sample: LC044

Location Code: LC044

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 6:00

Lab Log# FI25051-05

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.0369 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:39 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:39 @PD

Sample: LC033

Location Code: LC033

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 9:00

Lab Log# FI25051-06

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.295 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:43 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:43 @PD

Sample: LC048

Location Code: LC048

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 8:30

Lab Log# FI25051-07

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.153 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:47 @PD
Lead (Pb) EPA 200.8	Lead	0.0012 mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:47 @PD

Sample: LC024

Location Code: LC024

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 6:45

Lab Log# FI25051-08

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.186 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:51 @PD
Lead (Pb) EPA 200.8	Lead	0.0020 mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:51 @PD

Sample: LC026

Location Code: LC026

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 6:00

Lab Log# FI25051-09

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.0831 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 01:56 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 01:56 @PD

Sample: LC031

Location Code: LC031

PWSID#: OK3006604

Collection Type: Grab

Sample Time: 9/25/23 7:15

Lab Log# FI25051-10

Method/Parameter	Test	Result	Notes	PQL#	Prep Info	Analysis Info
Copper (Cu) EPA 200.8	Copper	0.206 mg/L		0.0100	09/26/23 18:00 @PD	09/27/23 02:00 @PD
Lead (Pb) EPA 200.8	Lead	BPQL mg/L		0.0010	09/26/23 18:00 @PD	09/27/23 02:00 @PD

Notes and Definitions

MCL Analyte concentration may exceed Maximum Contaminant Limit (MCL) for EPA Primary or Secondary Drinking Water Regulations.

Analyte concentration may exceed regulatory limit.

PQL Practical Quantitation Limit - the method reporting limit (MRL) adjusted for any dilutions or other changes made to the sample to deal with interferences/matrix effects

BPQL Below Practical Quantitation Limit (if applicable).

The "Prep Date" of the QC analysis coincides with the characters of the appropriate QC Lab ID. (Example: 19 A 02 15 - BLK = 2019, Jan 2, Batch #15 - Blank)

Lab Manager

A handwritten signature in black ink on a light gray background. The signature appears to be "Dg Cu" written in a cursive style.

Quality Control Data

Blank Data

QC Lab #	Test Group	Test	Result	PQL	Flags
23I2622-BLK1	Copper (Cu) EPA 200.8	Copper	BPQL mg/L	0.0100	
23I2622-BLK1	Lead (Pb) EPA 200.8	Lead	BPQL mg/L	0.0010	

Laboratory Control Sample Data

Lab QC#	Test Group	Test Name	LCS Result	Spike Level	Units	% Rec.	Control Limits	Flags
23I2622-BS1	Copper (Cu) EPA 200.8	Copper	0.0992	0.1000	mg/L	99	90 - 110	
23I2622-BS1	Lead (Pb) EPA 200.8	Lead	0.104	0.1000	mg/L	104	90 - 110	
23I2622-MRL1	Copper (Cu) EPA 200.8	Copper	0.00369	0.005000	mg/L	74	50 - 150	
23I2622-MRL1	Lead (Pb) EPA 200.8	Lead	0.005280	0.005000	mg/L	106	50 - 150	


Matrix Spike Data

QC Lab #	Test Group	Test Name	Source Sample	Sample Result	Units	Spike Result	Spike Level	% Rec.	Acceptance Limits	Flags
23I2622-MS2	Copper (Cu) EPA 200.8	Copper	FI25051-01	0.0106	mg/L	0.105	0.1000	94	85 - 115	
23I2622-MS2	Lead (Pb) EPA 200.8	Lead	FI25051-01	BPQL	mg/L	0.102	0.1000	102	85 - 115	

Matrix Spike Duplicate Data

QC Lab #	Test Group	Test Name	Sample Result	Spike Result	Spike Level	Units	% Rec.	Rec. Limits	% RPD	RPD Limit	Flags
23I2622-MSD2	Copper (Cu) EPA 200.8	Copper	0.0106	0.104	0.1000	mg/L	93	85-115	0.9	20	
23I2622-MSD2	Lead (Pb) EPA 200.8	Lead	BPQL	0.101	0.1000	mg/L	101	85-115	0.8	20	

* Complete Entire COC to be in Compliance*



Chain of Custody

Rogers Co. RWD #7

SDWIS Analysis Monthly

Accurate Work Order #	Date Sample Taken	Time Sample Taken	Matrix or Source (Refer. Below)	Grab (G) or Com (C)	Client I.D. / Sample Location or DEQ / EPA Location Code	Field Results		Sample Preserv. & Container	Due Date
						(pH, Temp, Chlorine, ...)	(note analysis & units)		
F125051	9-25-23	6:30A	DW	G	LC030			1000mL Plastic	9/25/23
01		7:00A	DW	G	LC036			1000mL Plastic	
02		6:30A	DW	G	LC035			1000mL Plastic	
03		7:00A			LC047			1000mL Plastic	
04		6:00A			LC044			1000mL Plastic	
05		9:00A			LC033			1000mL Plastic	
06								1000mL Plastic	

On-Site Info: Raw Alkalinity (TOC Raw) = _____ mg/L (E.Coli) = _____
 Matrix Codes: DW = Drinkingwater; WW = Wastewater; SL = Sludge; O = Other
 E. Coli Source: FS = Flowing Stream; RL = Reservoir/Lake; GWUDI = Groundwater under direct influence of surface water

COC Date - 10/02/09

Certification by Company Official: I hereby certify that the above sampling occurred during a period such that the sample(s) is/are representative of a typical operating day discharge for the above facility.

Sampled By: Charles Tipton Company: Rogers Co. RWD #7

Relinquished By: Charles Tipton Date/Time: 9-25-23

Received at Lab By: Charles Tipton Date/Time: 9-25-23 13:50

Reporting Requirements (standard 10-15 working days) Compliance Reporting? (DMR, PWS, Yes or No) OKlahoma PWS ID #

Mail Report To: Rogers Co. RWD #7 20352 S 4230 Rd Claremore, OK 74019

Address: 20352 S 4230 Rd Claremore, OK 74019

Phone #: (918) 341-1115 Fax #: (918) 342-1007

Email: waterdistrict7@yahoo.com

WWW.ACCURATELABS.COM (800) 516-5227

505 South Lowry Street Stillwater, OK 74074 Phone: (405) 372-5300 Fax: (405) 372-5396

12036 N. Pennsylvania Oklahoma City, OK 73120 Phone: (405) 751-3132 Fax: (405) 751-3108

-- All Glass containers provided by Accurate Labs have Teflon lined lids --
 -- All samples are scheduled to be disposed of in 4 weeks of receipt at Accurate --
 -- Hazardous samples will be returned to client or will be disposed of for a fee --

Signature: Charles Tipton Date/Time: 9-25-23

Sample Method:

Received By: Charles Tipton Date/Time: 9-25-23

Received at Lab By: Charles Tipton Date/Time: 9-25-23 13:50

Reporting Requirements (standard 10-15 working days) Compliance Reporting? (DMR, PWS, Yes or No) OKlahoma PWS ID #

Mail Report To: Rogers Co. RWD #7 20352 S 4230 Rd Claremore, OK 74019

Address: 20352 S 4230 Rd Claremore, OK 74019

Phone #: (918) 341-1115 Fax #: (918) 342-1007

Email: waterdistrict7@yahoo.com

WWW.ACCURATELABS.COM (800) 516-5227

505 South Lowry Street Stillwater, OK 74074 Phone: (405) 372-5300 Fax: (405) 372-5396

12036 N. Pennsylvania Oklahoma City, OK 73120 Phone: (405) 751-3132 Fax: (405) 751-3108

* Complete Entire COC to be in Compliance*



Chain of Custody

Rogers Co. RWD #7

SDWIS Analysis Monthly

Client Name -

Project Name -

Matrix or Source (Refer. Below)

Grab (G) or Com (C)

Time Sample Taken

Date Sample Taken

Client I.D. / Sample Location or DEQ/EPA Location Code

Field Results (pH, Temp, Chlorine, ...)

Location Code

Ch₂

of Container

Analysis Requested

Sample Preserv. & Container

Due Date

Accurate Work Order #	Date Sample Taken	Time Sample Taken	Matrix or Source (Refer. Below)	Grab (G) or Com (C)	Client I.D. / Sample Location or DEQ/EPA Location Code	Field Results (pH, Temp, Chlorine, ...)	Location Code	Ch ₂	# of Container	Analysis Requested	Sample Preserv. & Container	Due Date
FI25051	9-25-23	8:30 A	DW	G	LC048				1	lead + copper	125-ml Sterile Nalgene 1000ml plastic	9/25/23
07	"	6:45 A	DW	G	LC024				1			
08	"	6:00 A	DW	G	LC026				1			
09	"	7:15 A	DW	G	LC031				1			
10												

On-Site Info: Raw Alkalinity (TOC Raw) = _____ mg/L (E.Coli) = _____ ntu
 DW = Drinking water; WW = Wastewater; SL = Sludge; O = Other
 FS = Flowing Stream; RL = Reservoir/Lake; GWUDI = Groundwater under direct influence of surface water

COC Date - 10/02/09

Certification by Company Official: I hereby certify that the above sampling occurred during a period such that the sample(s) is/are representative of a typical operating day discharge for the above facility.

Sampled By: Charles Tipton Company: RWD7 Signature: Charles Tipton Date/Time: 9-25-23

Relinquished By: Charles Tipton Date/Time: 9-25-23

Received By: Charles Tipton Date/Time: 9-25-23 13:50

Reporting Requirements (standard 10-15 working days) Compliance Reporting? (DMR, PWS, Yes or No) Oklahoma PWS ID # OK3006604

Mail Report To: Rogers Co. RWD #7

Address: 20352 S 4230 Rd Claremore, OK 74019

Phone #: (918) 341-1115 Fax #: (918) 342-1007

Email: waterdistrict7@yahoo.com

WWW.accurateinlabs.com (800) 516-5227

505 South Lowry Street Phone: (405) 372-5300 Fax: (405) 372-5396

6558 E. 40th Street Phone: (918) 663-5400 Fax: (918) 663-6300

12036 N. Pennsylvania Oklahoma City, OK 73120 Phone: (405) 751-3132 Fax: (405) 751-3108

-- All Glass containers provided by Accurate Labs have Teflon lined lids --
 -- All samples are scheduled to be disposed of in 4 weeks of receipt at Accurate --
 -- Hazardous samples will be returned to client or will be disposed of for a fee --

Sample Method:

RUSH Request (if available) (Working Days)

Bid # -

PO # -

Phone #: ()

Fax #: ()